

Test Report Rev. 1.0

Objective of test

To determine the IPCT transfer function dependence on
 a) the position of the conductor inside the sensor hole
 b) the angle of the conductor inside the sensor hole

Date of test:
 By:
 Location:
 Temperature:

December 13, 1995, 8h45-9h15
 Alain Charvet and Julien Bergoz
 Crozet workshop
 22.8 °C


Unit tested:

IPCT-100T serial #039
 Burden resistor $1k\Omega \pm 1\%$
 Corresponding full scale range: $\pm 1A$


Test conditions:
 Primary current source:

AWG20 isolated wire thru the IPCT sensor hole
 1A constant current power supply thru 10Ω

Test data



Position X	Output [V]	Equivalent current [A]	Error X axis	Position Y	Output [V]	Equivalent current [A]	Error Y axis
30	9,987	0,9987	-0,13%	30	9,981	0,9981	-0,19%
20	9,986	0,9986	-0,14%	20	9,980	0,9980	-0,20%
10	9,985	0,9985	-0,15%	10	9,980	0,9980	-0,20%
0	9,985	0,9985	-0,15%	0	9,980	0,9980	-0,20%
-10	9,984	0,9984	-0,16%	-10	9,979	0,9979	-0,21%
-20	9,984	0,9984	-0,16%	-20	9,978	0,9978	-0,22%
-30	9,983	0,9983	-0,17%	-30	9,978	0,9978	-0,22%



Position α	Output [V]	Equivalent current [A]	Error X axis
0	9,981	0,9981	-0,19%
10	9,980	0,9980	-0,20%
20	9,980	0,9980	-0,20%
30	9,979	0,9979	-0,21%
40	9,979	0,9979	-0,21%
50	9,978	0,9978	-0,22%

Primary current drift during tests			
Time	Event	Primary Current	Drift
8h55	Start	1,000	
9h15	End	0,999	-0,10%

Remark: The IPCT output drift appears to be due to the 1A power supply drift (cold start).

Measurement of Output voltage vs. Input current

Unit tested: IPCT-10T, serial #021, with 100 mA full scale range (1kΩ ±0.01% burden resistor)

Date: 6 september 1995


Temperature 22.6 °C

1st measurement: Voltage drop across a standard 1 kΩ+0.0006% resistor in the input current

2nd measurement: Output voltage from the IPCT

3rd measurement: Output voltage from the IPCT when input current is disconnected

IPCT Range (mA)	Input voltage (V)	Precision resistor (Ω)	Equ. Input current (mA)	Output voltage (V)	Offset voltage (mV)	Adjusted output (V)	Equivalent current (mA)	Abolute error (μA)	Relative error (% of FS)
100	10,038	100	100,380	10,049	1,130	10,048	100,4767	-96,7000	-0,097%
100	8,051	100	80,510	8,060	1,060	8,059	80,5894	-79,4000	-0,079%
100	6,088	100	60,880	6,096	1,060	6,095	60,9494	-69,4000	-0,069%
100	4,049	100	40,490	4,057	1,070	4,056	40,5593	-69,3000	-0,069%
100	2,048	100	20,480	2,051	1,141	2,050	20,4986	-18,5900	-0,019%
100	10,016	1000	10,016	1,004	1,600	1,002	10,0240	-8,0000	-0,008%
100	8,090	1000	8,090	0,806	1,700	0,804	8,0430	47,0000	0,047%
100	6,012	1000	6,012	0,604	2,200	0,602	6,0180	-6,0000	-0,006%
100	4,031	1000	4,031	0,406	2,200	0,404	4,0380	-7,0000	-0,007%
100	2,028	1000	2,028	0,205	2,200	0,203	2,0280	0,0000	0,000%
100	1,005	1000	1,005	0,103	3,200	0,100	0,9980	7,0000	0,007%
100	0,803	1000	0,803	0,0835	3,120	0,080	0,8038	-0,8000	-0,001%
100	0,601	1000	0,601	0,0634	3,400	0,060	0,6000	1,0000	0,001%
100	0,400	1000	0,400	0,0436	3,400	0,040	0,4020	-2,0000	-0,002%
100	0,202	1000	0,202	0,0236	3,300	0,020	0,2030	-1,0000	-0,001%
100	0,101	1000	0,101	0,0135	3,500	0,010	0,1000	0,5000	0,001%
100	0,0801	1000	0,080	0,0115	3,400	0,008	0,0810	-0,9000	-0,001%
100	0,0605	1000	0,060	0,0096	3,500	0,006	0,0610	-0,5000	-0,001%
100	0,0404	1000	0,040	0,0077	3,600	0,004	0,0410	-0,6000	-0,001%
100	0,0200	1000	0,020	0,0056	3,600	0,002	0,0200	0,0000	0,000%



Output voltage vs. input current


Unit tested: IPCT-10T, serial #021, Range selected: 100 mA ÷ 10 V

Date: 5 september 1995

Test instrument: HP 8751A Baseband network analyzer

Stimulus signal +10 dBm thru HP 87512A Transmission/Reflexion test set. Attenuation -10dB, termination in 50Ω.

Signal from IPCT-10T into Network analyzer thru an series 1kΩ resistor (for input protection, negligible effect on measurement)




START 5 Hz

STOP 15 kHz

N STIMULUS val

1	5 Hz	-20.696 dB
2	100 Hz	-20.475 dB
3	1 kHz	-20.504 dB

Phase shift vs. frequency



START 5 Hz

STOP 15 kHz

N STIMULUS val

1	5 Hz	-1.5016 °
2	100 Hz	-912.93 m °
3	1 kHz	-20.898 °

Noise and residual modulator ripple in the IPCT-10T output signal

Unit tested: IPCT-10T, serial #056, Range selected: 10 mA ÷ 10 V

Date: March 7, 1996

Test instruments:


Function generator Philips PM 5132

Ouput signal into a 50Ω calibrated resistor

Oscilloscope HP 54503A with HP plotter

Current passing thru the IPCT sensor is adjusted to about. ±10mA


Plot of the IPCT-10T full swing signal: square wave +10mA...-10mA



Main	Timebase	Delay/Pos	Reference	
	20.0 ms/div	0.00000 s	Center	
Channel 1	Sensitivity	Offset	Probe	Coupling
	5.00 V/div	0.00000 V	1.000 : 1	dc (1M ohm)
Channel 2	5.00 V/div	0.00000 V	1.000 : 1	dc (1M ohm)

Trigger mode : Edge
 On Positive Edge Of Chan2
 Trigger Level
 Chan2 = 0.00000 V (noise reject ON)
 Holdoff = 40.000 ns

Zero noise magnification: IPCT-10T in ±10mA range



Main	Timebase	Delay/Pos	Reference	
	50.0 us/div	0.00000 s	Center	
Channel 2	Sensitivity	Offset	Probe	Coupling
	20.0 mV/div	-80.000 mV	1.000 : 1	dc (1M ohm)

Trigger mode : Edge
 On Positive Edge Of Chan2
 Trigger Level
 Chan2 = -50.400 mV (noise reject OFF)
 Holdoff = 40.000 ns

In range ±10mA ÷ ±10V:

Noise ≈ 20 mV rms in ±10V FS

Rms noise ≈ 0.2 % FS